


## LPNET07



### LPNET07 NET IRRADIANCE METER

LPNET07 measures the net radiation across a surface, from near ultraviolet to far infrared. The Net radiation is defined as the difference between the radiation that reaches the upper surface and the irradiation on the lower surface of the net radiometer. The surface of the upper receiver measures the direct solar radiation plus the diffuse one and the radiation at longer wavelengths emitted from the sky (clouds), while the lower receiving area measures the solar radiation reflected from the ground (albedo) and the radiation length wavelengths emitted from the earth.

The instrument is designed and constructed to be used outdoors in any weather conditions.

Besides its use in meteorology to measure energy balance, the LPNET07 can be used indoors for the measurement of radiant temperature (ISO 7726




FIG. 1 - Wiring diagram.

| Technical specifications |   |
|--------------------------|---|
| Typical Sensitivity      | 10 $\mu\text{V}/(\text{W}/\text{m}^2)$                      |
| Impedance                | $2 \Omega \div 4 \Omega$                                    |
| Measuring range          | $\pm 2000 \text{ W}/\text{m}^2$                             |
| Spectral range:          | $0.2 \mu\text{m} \div 100 \mu\text{m}$                      |
| Working temperature      | $-40 \text{ }^\circ\text{C} \div 80 \text{ }^\circ\text{C}$ |
| Weight                   | 0.35 kg   |
| Response time (95%)      | <60 s   |
| Field of view            | 180° upper sensor<br>180° lower sensor                      |


### Working Principle

The net radiometer LPNET07 is based on a thermopile sensor whose warm joints are in thermal contact with the receiver while the upper cool joints are in thermal contact with the lower receiver. The temperature difference between the two receivers is proportional to the net irradiance. The temperature difference between hot and cold junction is converted into a voltage by Seebeck effect. The two receivers are made from a portion of spherical coated PTFE. The particular form of the two receivers provides a response in accordance with the cosine. The PTFE coating, as well as allowing outdoor installation for long periods without risk of damage, can have a constant spectral response from ultraviolet (200nm) up to far infrared (100  $\mu\text{m}$ ).

### Installing and mounting the net radiometer for total irradiance measurements:

- To allow cleaning the two receiving surfaces regularly, LPNET07 should be mounted in easily reachable places. The surfaces can be washed with plain water or pure ETHIL alcohol.
- Mount the instrument so that no shadow will be cast on it at any time of day and of the seasons, from obstructions such as buildings, trees, or any other obstacle.
- In the NORTHERN hemisphere, the net radiometer is normally oriented towards SOUTH, while it should be oriented NORTHWARD, in the SOUTHERN hemisphere.
- The instrument should be mounted at a height of at least 1.5 m above the ground. Please note that the flow on the lower receiver is representative of a circular area with a radius of 10 times the height.
- When installing the net-radiometer avoid, wherever possible, to touch the surfaces of the receiving net-radiometer.

### Connection Diagram LPNET07



### Electrical Connections and requirements for electronic reading:

- LPNET07 does not require any power supply.
- It is available with a 5 m output cable
- It is supplied with a PTFE, UV resistant, braided shield and 2-wire cable. The colour code is as follows:  
black (shield) —> connected to the housing  
red —> (+) positive pole of the signal generated by the detector  
blue —> (-) negative pole of the signal generated by the detector
- It has to be connected to a millivoltmeter or to a data acquisition system with input impedance higher than 4000  $\Omega$ . Normally, the output signal from the net radiometer does not exceed  $\pm 20 \text{ mV}$ . In order to grant the best performances in measurements, the instrument resolution should be of  $1 \mu\text{V}$ .

## Maintenance:

In order to ensure a high measurement accuracy, it is necessary to keep the two receiving surfaces clean, the higher the frequency of cleaning, the best measurement accuracy will be.

Cleaning can be done with normal tissue for the cleaning of lens and water, if not enough, just use pure ethyl alcohol. After cleaning with alcohol it is necessary to clean the domes again with water only.

We strongly recommend to calibrate LPNET07 annually. The calibration can be carried out by comparison with another net-radiometer sample in the field. The field calibration is less precise than a calibration performed in the laboratory but has the advantage of not having to remove the instrument from its housing.

## Calibration and measurements:

Net radiometer sensitivity, indicated as  $S$  (or calibration factor), allows determining the net radiant flow passing through a surface. **S factor is measured in  $\mu\text{V}/(\text{Wm}^{-2})$ .**

Measured the potential difference (DDP) at the ends of the flow probe is obtained by the following formula  $E_e$

$$E_e = \text{DDP}/S$$

where;

$E_e$ : indicates the radiant flux expressed in  $\text{W}/\text{m}^2$ ,

DDP: indicates the potential difference expressed in  $\mu\text{V}$  and measured by the multimeter,

$S$ : indicates the calibration factor expressed in  $\mu\text{V}/(\text{W}/\text{m}^2)$  and shown on the net radiometer label (calibration factor is also mentioned in the calibration report).

N.B. If the difference of potential (DDP) is positive, the radiation on the upper surface is higher than the radiation on the lower surface (typically during daylight hours); if DDP is negative, the radiation on the lower surface is higher than the one on the upper surface (typically at night).

Each net-radiometer is individually calibrated at the factory and is distinguished by its calibrator factor.

Calibration is performed inside Delta OHM Metrological Laboratory and performed with a net radiometer-reference with a solar simulator as the source of light. Calibration is performed with a beam of light in parallel.

## Sensitivity to wind speed:

At the same radiant flux, by increasing the wind speed decreases the net radiometer output signal will (sensitivity decrease by increasing wind speed).

Measurements taken inside the wind tunnel, have shown that  $S_v$  sensitivity, related to the wind speed for LPNET07, can be corrected by using the following functions:

$$S_v = S_0(1-0.011 \times V) \quad \text{per } V \leq 10 \text{ m/s}$$

$$S_v = S_0(0.95-0.006 \times V) \quad \text{per } 10 \text{ m/s} < V < 20 \text{ m/s}$$

Where:  $S_0$  = sensitivity at zero wind speed  
 $V$  = wind speed in m/s




Fig. 2 - Calibration factor related with wind speed

Once we know both the net radiation - calculated through the sensitivity at zero wind speed ( $F_{\text{net}_0}$ ) - and the wind speed in ( $V$ ) in m/s, the correct data is obtained by using the following formula:

$$F_{\text{net}} = F_{\text{net}_0} / (1 - 0.011 \times V) \quad \text{per } V \leq 10 \text{ m/s}$$

$$F_{\text{net}} = F_{\text{net}_0} / (0.95 - 0.006 \times V) \quad \text{per } 10 \text{ m/s} < V < 20 \text{ m/s}$$

## Cosine response/Directional error:

The radiation falling on a surface should be measured with a sensor, whose response related to the light incidence angle, has to be a Lambertian Response. A receiver is known as Lambertian when its sensibility ( $S_\theta$ ), related to the incidence angle between the light and the detector surface, has the following behavior:

$$S_\theta = S_0 \cos(\theta)$$

Where:  $S_0$  is the sensitivity when light strikes perpendicular to the surface,  $\theta$  is the angle between the incident light beam and the line which is normal to the surface.




Fig. 3 - Typical behaviour of the error related to the angle of incidence.

## ORDERING CODES:

LPNET07: Net radiometer. Connecting cable: 5 m standard length, complete with  $\varnothing 16 \times 500$  rod for attachment to a mast Different cable lengths upon request.

HD978TR3: Configurable signal converter amplifier with 4...20 mA (20...4 mA) output. Input measuring range -10...+60 mV. Default setting 0...20 mV. 2-module DIN housing. Minimum measuring range 2 mV.

HD978TR5: Wall configurable, signal converter amplifier with 4...20 mA (20...4 mA) output. Input measuring range -10...+60 mV. Default settings 0...20 mV. Minimum measuring range 2mV. Wall mouting.

HD978TR4: Configurable signal converter amplifier with 0...10 Vdc (10...0 Vdc) output. Input measuring range -10...+60 mV. Default setting 0...20 mV. Minimum measuring range 2 mV. 2-module DIN housing.

HD978TR6: Wall configurable, signal converter amplifier with 0...10 Vdc (10...0 Vdc) output. Input measuring range -10...+60 mV. Default settings 0...20 mV. Minimum measuring range 2 mV. Wall mouting.



**DICHIARAZIONE DI CONFORMITÀ UE  
EU DECLARATION OF CONFORMITY**

**Delta Ohm S.r.L. a socio unico – Via Marconi 5 – 35030 Caselle di Selvazzano – Padova – ITALY**

Documento Nr. / Mese.Anno: **5170 / 07.2019**  
Document-No. / Month.Year :

Si dichiara con la presente, in qualità di produttore e sotto la propria responsabilità esclusiva, che i seguenti prodotti sono conformi ai requisiti di protezione definiti nelle direttive del Consiglio Europeo:  
*We declare as manufacturer herewith under our sole responsibility that the following products are in compliance with the protection requirements defined in the European Council directives:*

Codice prodotto: **LPNET07**  
Product identifier :

Descrizione prodotto: **Radiometro netto**  
Product description : **Net radiometer**

I prodotti sono conformi alle seguenti Direttive Europee:  
*The products conform to following European Directives:*

| Direttive / Directives   |  |
|--------------------------|--|
| 2014/30/EU               | Direttiva EMC / EMC Directive                    |
| 2014/35/EU               | Direttiva bassa tensione / Low Voltage Directive |
| 2011/65/EU - 2015/863/EU | RoHS / RoHS                                      |

Norme armonizzate applicate o riferimento a specifiche tecniche:  
*Applied harmonized standards or mentioned technical specifications:*

| Norme armonizzate / Harmonized standards |   |
|--|---|
| EN 61010-1:2010                          | Requisiti di sicurezza elettrica / Electrical safety requirements |
| EN 61326-1:2013                          | Requisiti EMC / EMC requirements                                  |
| EN 50581:2012                            | RoHS / RoHS   |

Il produttore è responsabile per la dichiarazione rilasciata da:  
*The manufacturer is responsible for the declaration released by:*

Johannes Overhues

Amministratore delegato  
Chief Executive Officer

Caselle di Selvazzano, 19/07/2019

Questa dichiarazione certifica l'accordo con la legislazione armonizzata menzionata, non costituisce tuttavia garanzia delle caratteristiche.  
*This declaration certifies the agreement with the harmonization legislation mentioned, contained however no warranty of characteristics.*

## WARRANTY


Delta OHM is required to respond to the "factory warranty" only in those cases provided by Legislative Decree 6 September 2005 - n. 206. Each instrument is sold after rigorous inspections; if any manufacturing defect is found, it is necessary to contact the distributor where the instrument was purchased from. During the warranty period (24 months from the date of invoice) any manufacturing defects found will be repaired free of charge. Misuse, wear, neglect, lack or inefficient maintenance as well as theft and damage during transport are excluded. Warranty does not apply if changes, tampering or unauthorized repairs are made on the product. Solutions, probes, electrodes and microphones are not guaranteed as the improper use, even for a few minutes, may cause irreparable damages.

Delta OHM repairs the products that show defects of construction in accordance with the terms and conditions of warranty included in the manual of the product. For any dispute, the competent court is the Court of Padua. The Italian law and the "Convention on Contracts for the International Sales of Goods" apply.

## TECHNICAL INFORMATION

The quality level of our instruments is the result of the continuous product development. This may lead to differences between the information reported in the manual and the instrument you have purchased. In case of discrepancies and/or inconsistencies, please write to sales@deltaohm.com.

Delta OHM reserves the right to change technical specifications and dimensions to fit the product requirements without prior notice.



## DISPOSAL INFORMATION

Electrical and electronic equipment marked with specific symbol in compliance with 2012/19/EU Directive must be disposed of separately from household waste. European users can hand them over to the dealer or to the manufacturer when purchasing a new electrical and electronic equipment, or to a WEEE collection point designated by local authorities. Illegal disposal is punished by law. Disposing of electrical and electronic equipment separately from normal waste helps to preserve natural resources and allows materials to be recycled in an environmentally friendly way without risks to human health.



**RoHS**